

# Observations after 15 years of Platelet Bacterial Screening at the Irish Blood Transfusion Service

Danny Curran


[Danny.curran@ibts.ie](mailto:Danny.curran@ibts.ie)

3<sup>rd</sup> October 2018

# Timeline

- Nov 2004 Platelet Screening
- Sep 2005 BPN Bottle
- Nov 2005 Day 4 Testing
- July 2007 Orbisac Pools
- March 2015 Testing at the NBC centre only

# BacT Testing



# Sampling Protocol

- Apheresis sampled day after donation (day 1)
- Pools sampled day after pooling (day 2)
- BPA & BPN bottles
- Incubated to beyond day 5 expiry  
(Aph 4.8 days or 3.8 days Pools)
- Product released after loading verification
- Additional sample day 4 (for 7 day shelf life)
- Loading two batches per day (Mon to Saturday)

# Numbers

<b>Overall Program</b>	<b># Tested</b>	<b>false rate</b>	<b>Non- confirmed rate</b>	<b>confirmed rate</b>
<b>Apheresis</b>	122,311	0.10	0.07	0.04
<b>Pools</b>	76,940	0.37	0.10	0.13

# Bacterial Screening Test Results **Apheresis Platelets**

## 2005-2018 (August)

Apheresis Platelets 2005* - 2018	Donations	Non-confirmed Positives	Non-confirmed Positive Rate %	Confirmed Positives	Confirmed Positive Rate %	Total Positive Rate (confirmed plus non-confirmed) %
Day 1 (Initial) Test	122311	55	0.045	35	0.029	0.073
Day 4 Test	58709**	16	0.027	7	0.012	0.039
Expired (tested on day 8)	2374	1 <sup>#</sup>	0.042	0	0	0.042
Overall	-	<b>72</b>	<b>0.114</b>	<b>42</b>	<b>0.041</b>	
<b>Overall Positive rate</b>						<b>0.154% (1:649)</b>
*Post use of BPA & BPN bottles **Estimated Quantity of donations retested at day 4 (based on a 2016 rate of 48% retested) # 2011- re-testing of Product was negative						


# Bacterial Screening Test Results **Pooled** Platelets

## 2007-2018 (August)

Pooled Platelets 2007*- 2018	Total	Non-confirmed Positives	Non-confirmed Rate %	Confirmed	Confirmed Rate %	Total Positive Rate (confirmed plus unconfirmed) %
<b>Day 2</b> (Initial)	76940	42	0.055	65	0.084	<b>0.139</b>
<b>Day 4</b>	20615	9	0.044	9	0.044	<b>0.088</b>
<b>Expired</b> (tested on day 8)	694	0	0.000	0	0.000	<b>0.000</b>
Overall	-	51	<b>0.099</b>	74	<b>0.128</b>	
<b>Overall Positive rate</b>						<b>0.227 % (1:441)</b>
* Post start of Orbisac Pool Platelets						

# Bacteria isolated


**Confirmed positive platelets (116)**





# Bacteria: Non-confirmed Positives

- 122 Non-confirmed Positives
- 66 had a platelet re-tested
- 15 Bacillus (0 Bacillus with confirmed positives).




# Day 4 testing Analysis

Initial V day 4 Confirmed Positive rates:

	Initial test Confirmed rate %	Day 4 Test Confirmed rate %	Initial Test Efficacy
<b>Apheresis</b>	0.029	0.012	70.7%
<b>Pools</b>	0.084	0.044	65.6%

# Day 4 test Bacteria

	% Propionibacterium spp.	
	Day 2	Day 4
Pool Platelets	46.2% (30 out of 65)	33.3% (3 out of 9)
Apheresis Platelets	52.9% (18 out of 34)	14.3% (1 out of 7)



# Positive Investigation

## Retesting success

- Detection success of positive bacteria in associated Red cells?

Bacteria detected in Pool Platelet	# occurrences 4 FRCC Tested	# where FRCC positive	% Recovery in FRCC
Propionibacterium	29	28	96.6%
Staphylococci	27	4	14.8%
Streptococci	7	1	14.3%

- Pool only positive (29 events) – 27 Staphylococci/Streptococci.
- Red cell only positive (7 events) - 5 Propionibacterium.
- **Summary:** Gram positive cocci are very less likely to be detected in Red cell packs.

# Retesting Product


## (Influence of pack Incubation boost)

- 2012 started incubating suspected positive packs at 32°C in tandem with the retest.
- 33% (10 out of 30) initial negative packs then gave a positive result in a further test post incubation.
- E.g. May 2018: Positive pool (initial day 2 sample), P.acnes.  
Day 4 sample negative, Pool initial retest negative & FRCC positive.  
Pool 2<sup>nd</sup> retest post incubating positive!
- Summary: a portion of non-confirmed results will be confirmed by further testing after incubation of packs.

# Recall success


- Confirmed positives = 71% Recall success (34/116 platelets infused).

Bacteria infused:



- Total (Conf + Non- Conf) = 63% Recalled (89/238 infused).
- No platelet Septic TTI reported.

# False positive history



# PAS influence

	# Products	# Positives	Positive Rate %
<b>Pre-PAS</b> (2004-July 2007)	29,352	12	0.041
<b>Post PAS</b>	76,940	65	0.096

- TTD Apheresis V PAS Pools – no obvious difference noted
- Had a big influence on false positives



# Other learnings

- 2011: Controller fault.
- 2015: drawer/rack faults
- 2017 Power interruption..plan contingency
- 2018: Rise in non-confirmed results following disinfectant change - bottle septum disinfection critical

# Testing Differences

- Sampling device (pouch/device)
- Sampling method – mother bag/each split
- Sampling time
- One/both bottles
- Incubation time
- Hold time
- Dealing with positives/product recalls
- Testing multiple batches/different centers
- Terminology

= Difficult to directly compare results across different organisations

# Future

- Later Apheresis sampling?
- Single test?
- Hold period?
- Link to eprogesa?
- Contingency?
- Pathogen reduction

# Thanks to:

- Michael Maher, Senior Medical Scientist
- Dr Niamh O'Flaherty, Consultant Microbiologist