



A UKAS EXPERIENCE

JAMES TAYLOR
TRANSFUSION LABORATORY MANAGER
HEART OF ENGLAND FOUNDATION TRUST

NEQAS/BBTS SIG
10.11.15



UKAS

**UNITED
KINGDOM
ACCREDITATION
SERVICE**

QUALITY MANAGEMENT SYSTEM (QMS)

- ✘ Audit
- ✘ Training
- ✘ Meeting minutes
- ✘ Suppliers
- ✘ Improvement Action Report (IAR)
- ✘ IQC/EQA

TRACEABILITY AND UNCERTAINTY OF RESULTS

- ✘ “what on earth is that?”
- ✘ Isn't it just for Biochemistry?

- ✘ Pre analytical
- ✘ Analytical
- ✘ Post analytical

- ✘ Metrological not meteorological!!

IMPACT ON THE WORLD OF TRANSFUSION

Calibration issues

- ✘ Problems for Diamed users

- ✘ Problems for others?






IMPACT ON THE WORLD OF TRANSFUSION

✘ Temperature mapping

For fridges is it 2 – 6°C or is it
2.5 – 5.5°C?







UKAS

**UNITED
KINGDOM
ACCREDITATION
SERVICE**

**ISO
15189**