Safeguarding public health



## **REGULATORY HAEMOVIGILANCE:**

## The root causes of 'human error'

Judy Langham Principal Haemovigilance Specialist

© Crown copyright 2005



# Serious adverse blood reactions and events (SABRE)

- SABRE data 2005 2012
- Serious adverse events EU Directive ' Deviations and Specifications'
- Outcomes versus root causes
- Why do root cause analysis?
- Case study



## Why undertake haemovigilance? What is haemovigilance?

 a set of organised surveillance procedures relating to serious adverse or unexpected events or reactions in donors or recipients (of blood and blood components) and the epidemiological follow-up of donors

#### Why do we do it?

- to discharge our legal obligations
- to inform the European Commission
- to inform the Department of Health, National Transfusion Committee and SaBTO

• to advise on improvements to the safety of transfusion practice based on the data supplied by reporters



# Why have an incident management system?

• to prompt any follow-up action necessary to protect other patients from suspected hazards e.g bacterial contamination

• to support the development of clinical guidelines for hospitals in relation to the use of blood and blood components

• to support the training of medical, nursing and technical staff


• to provide summary data over a period of time which may highlight emerging trends

• to use as evidence to drive process improvements



### All UK SABRE reports by year

Excluded SAR SAE Total



UK NEQAS/ BBTS Blood Bank technology SIG Judy Langham, MHRA



#### **All UK Serious Adverse Events 2011**

Deviation /specification	Total Number	Product Defect	Equipment Failure	Human Error	Other
Whole blood collection	36	0	0	36	0
Apheresis collection	1	0	0	1	0
Testing of donations	8	0	0	8	0
Processing	37	3	1	32	1
Storage	228	1	3	224	0
Distribution	52	0	0	51	1
Materials	0	0	0	0	0
Other	449	3	10	436	0
Overall Total:	811	7	14	788	2

Safeguarding public health



#### SABRE data 2012:

Closed reports (unverified) to Oct 31st

Total	=	1238	
Excluded	=	153	
SAR	=	306	
SAE	=	779	



#### **All UK Serious Adverse Events 2012**

Total Number	Product Defect	Equipment Failure	Human Error	Other
47	23	0	23	1
11	8	0	3	0
8	0	2	5	1
20	0	0	19	1
182	1	5	171	5
40	0	0	38	2
3	0	1	2	0
468	2	4	457	5
779	34	12	718	15
	47 11 8 20 182 40 3 468	47       23         11       8         8       0         20       0         182       1         40       0         3       0         468       2	47       23       0         11       8       0         8       0       2         20       0       0         182       1       5         40       0       0         3       0       1         468       2       4	4723023472302311803802520001918215171400038301246824457

Safeguarding public health Serious Adverse Events 2012:	MHRA
Other/Human Error Outcomes (unverified)	
<ul> <li>Incorrect blood component issued (missing special requirements)</li> </ul>	97
<ul> <li>Component labelling errors (at point of issued from lab)</li> </ul>	63
Sample processing errors	59
Data entry errors (lab IT)	58
Pre transfusion testing errors	55
Components available for transfusion past de-reservation date	40
Component collection errors	23
Failed recalls	8
Expired components available for transfusion	6




#### Root Causes of SAEs – human error (2011 n = 788)

Error Type	Definition
Incorrect process	Process does not achieve the desired outcome
Incorrect procedure	Written procedure does not reflect the process
Procedural steps omitted	Procedural steps missed out (Intentional or forgotten) -may be a result of rushing/concentration lapse
Lapsed/ no training	Training/competency assessment out of date, not completed
Inadequate training	Training/competency assessment does not cover error made
Ineffective training	Training is adequate, but has been misunderstood
Rushing	Working too quickly, failing to check for accuracy
Concentration	Error when not obviously omitting steps or rushing
Communication	Written/verbal communication not clear/inaccurate



### Root causes of SAEs (2011) specification Human Error (n = 788)



© Crown copyright 2008



#### **Targeted corrective measures**

#### **Procedural based errors:**

- Inadequate process
  - review existing system for process validation
- Incorrect procedure
  - reassess document control process
- Procedural steps omitted

 re-emphasise the importance of strict adherence to written protocols (a basic principle of Good Manufacturing Practice)



#### **Targeted corrective measures**

#### **Training based errors:**

- Lapsed/ no training review training and induction schedules and ensure these identify re-joiners
- Inadequate training reassess training material to ensure it covers all essential activities
- Ineffective training encourage staff to take responsibility for their own learning and to highlight any areas they have not fully understood



#### **Targeted corrective measures**

#### **Concentration and Communication:**

• Communication – encourage teamwork and the use of clear, unambiguous written and verbal communication protocols

• Rushing – encourage staff to work at an appropriate pace which allows them time to follow procedures precisely, prioritise workload effectively and self-check for accuracy

 Concentration – minimise distractions, advise staff to restart a process when they have been distracted and encourage individual reflective practice to understand what they would do differently in future



#### **Case Study**

#### **Notification: SAE/ Storage/ Human error**

• Blood was returned from the ward because the expiry date on the accompanying paperwork was incorrect

- Blood was out of the fridge for 33 minutes
- Lab staff amended the expiry date and returned the unit to stock

• This unit was then issued and transfused to another patient – unit was out of the fridge for a total of 49 mins (within the 4hr limit)



#### **Case Study**

#### **Confirmation:**

• Root cause = not following laboratory procedure

 Corrective measures = lab staff have been retrained in the correct procedures for booking in stock and on the importance of the cold chain. The 30 minute rule is now specifically discussed at GMP updates.

#### **COMMENTS**?

## Safeguarding public health Case Study

#### Suggestions :

- Report as SAE/ Other / human error (data entry error)
- Investigate WHY laboratory procedures were not followed
  i) why was the expiry date entered incorrectly?
  ii) why was the unit returned to stock?
- Root causes are likely to be

   rushing/ cutting corners by not scanning all units into stock individually
   lack of concentration when checking details prior to issue
   distraction when returning unit to stock rather than discarding as out of
   temperature control

#### **Corrective Measures:**

- Check patient receiving the out of temperature unit suffered no consequences
- Check lab staff are aware of the correct procedures if not then provide training
- Encourage lab staff to consider what they would do differently in similar circumstances

